

**DECLARATION OF CONFORMITY  
TO COUNCIL DIRECTIVE 93/42/EEC  
CONCERNING MEDICAL DEVICES**



**MANUFACTURER:**

**CONTEC MEDICAL SYSTEMS CO., LTD**  
No.24 Huanghe West Road Economic & Technical  
Development Zone ,Qinhuangdao,Hebei Province,  
066004,P.R.China

**MEDICAL DEVICE:**

Pulse Oximeter SAT500

**CLASSIFICATION - ANNEX IX:**

Class II b, Rule 10

**CONFORMITY ASSESSMENT ROUTE:** Annex II without chapter 4

WE, ( CONTEC MEDICAL SYSTEMS CO., LTD ) HEREWITH DECLARE THAT THE STATED  
MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF  
COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES;  
INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC  
ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH  
DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

**NOTIFIED BODY:**

TÜV SÜD PRODUCT SERVICE GMBH  
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

**IDENTIFICATION NUMBER:**

 0123

**(EC) CERTIFICATE(S):**

G1 13 06 50972 019

**EC REP**

**EUROPEAN REPRESENTATIVE:**


Shanghai International Trading Corp. GmbH(Hamburg)  
Eiffestrasse 80, 20537 Hamburg Germany

**START OF CE-MARKING:**

2006-09-28 (Date or Lot or serial number)

**PLACE, DATE OF DECLARATION:**

**SIGNATURE:**



\_\_\_\_\_  
President

# DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

## Appendix: list of (harmonised - EN) standards

NO.	Reference	Title
1	EN60601-1:1990+A1:1993+A2:1995 (IEC60601-1:1988+A1:1991+A2:1995)	Medical electrical equipment- Part 1: General requirements for safety
2	EN 60601-1-2:2007 (IEC60601-1-2:2007)	Medical electrical equipment- Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
3	EN60601-1-4:1996+A1:1999 (IEC60601-1-4:1996/A1:1999)	Medical electrical equipment - Part 1-4: General requirements for safety - Collateral standard: Programmable electrical medical systems
4	EN 60601-1-6:2007 (IEC60601-1-6:2006)	Medical electrical equipment-Part 1-6:General requirements for basic safety and essential performance-Collateral Standard: Usability
5	EN ISO 9919:2009 (ISO 9919:2005)	Medical electrical equipment - Particular requirements for the basic safety and essential performance of pulse oximeter equipment for medical use
6	EN 62304:2006	Medical device software-Software life-cycle processes