

declaration of conformity

We, the manufacturer **seca gmbh & co. kg.**, declare under our responsibility that the products of the category **non-automatic electrical baby infant scales**:

Products:	375	376	378	717	757	336
EC type approval	D05-09-024		D93-09-153		T7945	
Accuracy class	III		III		III	
Optional wireless transmission	x					
Classification medical devices	class I with measuring function					

meet all the provisions of the following directives which apply to it:

2009/23/EC Directive relating to non-automatic weighing instruments

93/42/ EEC and 2007/47/EC Directive concerning medical devices

Applied harmonised standards, national standards or other normative documents:

EN 45501 Metrological aspects of non-automatic weighing instruments

EN 60601-1 Medical electrical equipment - Part 1: General requirements for safety

EN 60601-1-2 Medical electrical equipment - Part 1-2: General requirements for safety
Collateral standard: Electromagnetic compatibility

Applicable for wireless transmission:

EN 300 328 Electromagnetic compatibility and Radio spectrum Matters 2,4GHz

EN 301489-1 Electromagnetic compatibility and Radio spectrum Matters
Part 1: Common technical requirements

EN 301489-17 Electromagnetic compatibility and Radio spectrum Matters
Part 17: Specific conditions for Broadband Data Transmission Systems

Notified Bodies: Hessische Eichdirektion TÜV SÜD Product Service GmbH
Holzhofallee 3 Ridlerstrasse 65
64283 Darmstadt 80339 München



Hamburg, 01 / 08 / 2011



Frederik Vogel
CEO Development & Manufacturing



Robert Vogel
CEO Sales & Marketing