

## declaration of conformity

We, the manufacturer seca gmbh & co. kg., declare under our sole responsibility that the products of the category non-automatic electrical chair scale:

| Products:                      | 955                             | 956 | 957 | 959 | 963 | 958 |
|--------------------------------|---------------------------------|-----|-----|-----|-----|-----|
| EC type approval               | D02-09-011                      |     |     |     |     |     |
| Accuracy class                 | III                             |     |     |     |     | Ш   |
| Classification medical devices | class I with measuring function |     |     |     |     |     |

## meet all the provisions of the following directives which apply to it:

**2009/23/EC** Directive relating to non-automatic weighing instruments

93/42/ECC and 2007/47/EC

Directive concerning medical devices

## Applied harmonised standards, national standards or other normative documents:

**EN 45501** Metrological aspects of non-automatic weighing instruments

EN 60601-1 Medical electrical equipment Part 1: General requirements for safety

**EN 60601-1-2** Medical electrical equipment Part 1-2: General requirements for safety

Collateral standard: Electromagnetic compatibility

Notified Bodies: Hessische Eichdirektion

Holzhofallee 3 64283 Darmstadt TÜV SÜD Product Service GmbH

Ridlerstrasse 65 80339 München



Hamburg, 01 / 06 / 2011

Frederik Vogel

CEO Development & Manufacturing

Robert Vogel

CEO Sales & Marketing